



**FILTERING FACEPIECE
APPROVAL LABEL**



**Masprot S.C.e.I. Ltda.
Carlos Walker Martinez 5574, San Miguel
Santiago, Chile
Tel: 0056-2-4989000**

THIS RESPIRATOR IS APPROVED ONLY IN THE FOLLOWING CONFIGURATION:

RESPIRATOR COMPONENTS			
TC-	PROTECTION¹	FACEPIECE	CAUTIONS AND LIMITATIONS²
		M300	
84A-5037	N95	X	A B C J M N O P S

1. PROTECTION

N95-Particulate Filter (95% filter efficiency level) effective against particulate aerosols free of oil; time use restrictions may apply.

2. CAUTIONS AND LIMITATIONS:

- A. Not for use in atmospheres containing less than 19.5 percent oxygen.
- B. Not for use in atmospheres immediately dangerous to life or health.
- C. Do not exceed maximum use concentrations established by regulatory standards.
- J. Failure to properly use and maintain this product could result in injury or death.
- M. All approved respirators shall be selected, fitted, used, and maintained in accordance with MSHA, OSHA and other applicable regulations.
- N. Never substitute, modify, add, or omit parts. Use only exact replacement parts in the configuration as specified by the manufacturer.
- O. Refer to user's instructions and/or maintenance manuals for information on use and maintenance of these respirators.
- P. NIOSH does not evaluate respirators for use as surgical masks.
- S. Special or critical User's Instructions and/or specific use limitations apply. Refer to User's Instructions before donning.